

SCOPE OF WORK

The provision of Calibration & repair services of various equipment for Eskom Distribution in the Cape Costal Cluster – Eastern Cape.

Description o of the Scope / Specifications

Calibration and minor repairs of various equipment, on an as when required basis, for a period of 3 years, within the Cape Costal Cluster – Eastern Cape operating unit.

This scope is applicable to equipment calibrated in Eskom premises and at the service provider laboratory.

The supplier shall calibrate the equipment the equipment as per the instruction from Eskom in the form of formal written letter (Task order) or official purchase order, therefore during the calibration should there be necessary need to replace parts, formal quotation must be sent to Eskom for acceptance or rejection.

In event that a quote is rejected an equipment will be return to Eskom with full calibration report

Upon completion of calibration of the equipment, the service provider shall ensure to paste a sticker on the equipment showing the following information:

- The calibration certificate must be issued together with the calibrated equipment.
- Defective report / certificate must be issued for all defective equipment, and should be the cost of repair be above the book value of the equipment, the equipment will be sent back to the end user with detailed feedback of the damages requiring repairs

Where equipment is returned to the End-user defective, a legible “defective” sticker as described below shall be attached to defective instruments by the accredited calibration / test facility before returning such equipment back to Ende-user

Requirements for sticker - it is applicable to both routine calibration sticker and defective equipment sticker.

The successful tenderers must be SANSI/ISO1725 Accredited

INSTRUMENT SHORT DESCRIPTION (ref BOQ)

Compiled by Officer Contracts - Name: Nokulunga Siguba

Signature

Approved by Technical Support Manager Name - Noxolo Mabula

Signature